

		code	BC30
Meetonzekerheid uit jaarrapport SKML		procedure	Carbamazepine (+OH Carbamazepine 10,11 epoxide)
	Criterion	Jaarscore	≥1
	T.E.a % SKML		20
Stof	Juistheid	%	-10,0
	Precisie	%	0,72
	Total error	%	11,44
	Sigma-TE		6,0
	Sigma-SA		6,0
	SKML Score		2
Meet onzekerheid	95%		1,4
	EIS		OK
	Beoordeling		OK
Metaboliet 1	Juistheid	%	-6,4
	Precisie	%	1
	Total error	%	8,4
	Sigma-TE		6,0
	Sigma SA		6,0
	SKML Score		2
Meet onzekerheid	95%		2,0
	EIS		OK
	Beoordeling		OK
Metaboliet 2	Juistheid	%	-7,4
	Precisie	%	0,58
	Total error	%	8,56
	Sigma-SA		
	Sigma-TE		6,0
	SKML Score		2
Meet onzekerheid	95%		1,1
	EIS		OK
	Beoordeling		OK
Stof	Bijdrage bias	%	87
	Bijdrage Imprecisie	%	13
Metaboliet 1	Bijdrage bias	%	76
	Bijdrage Imprecisie	%	
Metaboliet 2	Bijdrage bias	%	86
	Bijdrage Imprecisie	%	14