

# Performance data sheet

Testnaam	Kobalt/kreatinine [molaire ratio] in urine
LOINC code	34270-9
Detection limit	<4 nmol/l
Imprecision	Low 12,6 nmol/l: CV% 4,5 High 101,1 nmol/l: CV% 3,7
Measurement range	<4 - 910 nmol/l
Traceability	Atomic Spectroscopy Standard Cobalt 1000 mg/l in 2% HNO <sub>3</sub> , Perkin Elmer Interne standaard Ittrium, Perkin Elmer