

Performance data sheet

Testnaam	Koper [mol/volume] in serum of plasma
LOINC code	14665-4
Detection limit	0,27 umol/l
Imprecision	Low 12,0 umol/l: CV% 3,1 High 38,5 umol/l: CV% 3,1
Measurement range	0,27 - 113,5 umol/l
Traceability	Atomic Spectroscopy Standard Koper 1000 mg/l, Perkin Elmer Interne standaard Rhodium, Perkin Elmer