

Performance data sheet

Testnaam	Zink [mol/volume] in serum of plasma
LOINC code	14955-9
Detection limit	0,30 umol/l
Imprecision	Low 13,8 umol/l: CV% 2,9 High 42,5 umol/l: CV% 2,0
Measurement range	0,30 - 103,0 umol/l
Traceability	Atomic Spectroscopy Standard Zink 1000 mg/l in 2% HNO ₃ , Perkin Elmer Interne standaard Ittrium, Perkin Elmer