

Performance data sheet

Testnaam	Mangaan [mol/volume] in serum of plasma
LOINC code	25467-2
Detection limit	6,74 nmol/l
Imprecision	Low 13,2 nmol/l: CV% 7,8 High 119,9 nmol/l: CV% 4,1
Measurement range	6,74 - 941 nmol/l
Traceability	Atomic Spectroscopy Standard Mangaan 1000 mg/l, Perkin Elmer Interne standaard Ittrium, Perkin Elmer