

# Performance data sheet

**Testnaam** Sexhormoon bindend globuline [mol/volume] in serum of plasma

LOINC code 13967-5

Detection limit 2,0 nmol/l

Imprecision  
Low 1,2 nmol/l: CV% 4,2  
Mid 21 nmol/l: CV% 5,2  
High 80 nmol/l: CV% 6,6

Measurement range 2,0 - 180 nmol/l

Traceability Interne Standaard, Siemens