

Performance data sheet

Testnaam	Calcidiol [mol/volume] in serum of plasma
LOINC code	14635-7
Detection limit	10,0 nmol/l
Imprecision	Low 20,9 nmol/l: CV% 6,3 High 103,4 nmol/l: CV% 2,7
Measurement range	10,0 - 150,0 nmol/l
Traceability	The standards are prepared gravimetrically and are traceable to internal reference calibrator concentrations, Fujirebio