

# Performance data sheet

---

Testnaam	Anti Mullerse hormoon [massa/volume] in serum of plasma
----------	---

---

LOINC code	38476-8
Detection limit	0,1 ug/l
Imprecision	Low 2,8 ug/l: CV% 1,5 High 8,7 ug/l: CV% 1,8
Measurement range	0,1 - 25,0 ug/l
Traceability	Calibrator van Beckman Coulter