

| Mutatie | Techniek | Maximale gevoeligheid in kopieën / mL |
|-----------------------|--------------|---------------------------------------|
| EGFR L858R | ddPCR | 0,5 |
| EGFR Exon 19 deleties | ddPCR | 0,5 |
| EGFR L861Q | ddPCR | 0,5 |
| EGFR T790M | ddPCR | 0,5 |
| Mutaties Overig | ddPCR | 0,5 |