

Performance data sheet

Testnaam	Ascorbaat [mol/volume] in serum of plasma
LOINC code	14622-5
Detection limit	3,5 umol/l
Imprecision	Low 39,15 umol/l: CV % 9,0 High 119,02 umol/l: CV% 8,9
Measurement range	3,5 - 826 umol/l
Traceability	Grondstof: L(+)-Ascorbinezuur; Acros Organics; 99% zuiverheid; Casnr: 50-31-7.