

	code	CIPR
Meetonzekerheid uit jaarrapport SKML	Procedure	Citalopram (en Desmethylcitalopram, Di-desmethylcitalopram)
	Criterion	Jaarscore
	T.E.a % SKML	≥1
Stof	Juistheid	30
	Precisie	-2,8
	Total error	7,7
	Sigma-TE	18,2
	Sigma-SA	5,8
	SKML Score	3,0
Meet onzekerheid	95%	2
	EIS	15,1
	Beoordeling	OK
Metaboliet 1	Juistheid	OK
	Precisie	%
	Total error	%
	Sigma-TE	%
	Sigma SA	%
	SKML Score	%
Meet onzekerheid	95%	-12
	EIS	3,8
	Beoordeling	19,6
Metaboliet 2	Juistheid	6,0
	Precisie	3,3
	Total error	2
	Sigma-TE	7,4
	Sigma-SA	OK
	Sigma-TE	OK
	SKML Score	OK
Meet onzekerheid	95%	15
	EIS	85
	Beoordeling	61
Stof	Bijdrage bias	%
	Bijdrage Imprecisie	%
Metaboliet 1	Bijdrage bias	%
	Bijdrage Imprecisie	%
Metaboliet 2	Bijdrage bias	%
	Bijdrage Imprecisie	%